



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**
205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

FIRST QUADRIMESTER 2019

VAGINOSIS SCREEN

Candida albicans

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
BD Affirm VP111		18	18		18		17	1	18		18
Total Population	0	18	18	0	18	0	17	1	18	0	18
Flagging	***			***		***		***		***	

Gardnerella vaginalis

Method	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Labs
	BD Affirm VP111	18			18	18		18		18	
Total Population	18	0	0	18	18	0	18	0	18	0	18
Flagging		***	***			***		***		***	

Trichomonas vaginalis

Method	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Labs
	BD Affirm VP111		18	1	17	18		1	17	18	
Total Population	0	19	1	18	19	0	1	18	19	0	19
Flagging	***		***			***	***			***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.