



**Normal Mean (GMNPT)****Initial Grouping by Reagent and Instrument**

Siemens Innovin & Siemens/Sysmex CA series	1	9.2 - 11.2	S 10.21	0.32	-	-	-	-	-	-	-	-	-	-	-	-	36
<b>Initial Grouping by Reagent</b>																	
Siemens Innovin	2	9.3 - 11.1	S 10.19	0.31	-	-	-	-	-	-	-	-	-	-	-	-	40
<b>Initial Grouping by Sensitivity or Principle</b>																	
High sens/wide normal range	3	9.4 - 14.9	S 12.19	0.92	-	-	-	-	-	-	-	-	-	-	-	-	16
Other	4	7.9 - 14.4	S 11.13	1.09	-	-	-	-	-	-	-	-	-	-	-	-	37
High sens/mod normal range	5	9.3 - 11.1	S 10.19	0.31	-	-	-	-	-	-	-	-	-	-	-	-	40
<b>Total Population</b>																	
Whole Population	6	7.7 - 14.4	S 11.01	1.12	-	-	-	-	-	-	-	-	-	-	-	-	102

**Activated Partial Thromboplastin Time, Seconds****Initial Grouping by Reagent and Instrument**

HemosIL APTT-SP & IL ACL series 100 - 7000	1	32.2 - 43.5	P 37.83	1.36	26.0 - 35.1	P 30.54	1.54	35.6 - 48.1	P 41.85	2.33	38.5 - 52.1	P 45.29	1.11	23.8 - 32.2	P 28.02	1.81	15
HemosIL SynthASii w/CaCl2 & IL ACL Elite,E.Pro, 8-10000	2	33.9 - 45.9	P 39.88	1.31	33.6 - 45.5	P 39.56	2.3	30.3 - 41.0	P 35.62	1.25	42.8 - 57.9	P 50.37	1.11	23.9 - 32.3	P 28.1	0.94	12
Siemens FS, Actin Activated, FSL Activated & Siemens/Sysmex CA series	3	31.7 - 42.9	P 37.34	1.62	21.5 - 29.1	P 25.27	0.71	24.8 - 33.6	P 29.19	1.41	38.3 - 51.8	P 45.07	2.51	20.1 - 27.2	P 23.61	0.95	72
<b>Initial Grouping by Reagent</b>																	
Diag Stago STA-PTT-A	4	35.1 - 47.5	P 41.33	1.13	27.3 - 36.9	P 32.12	1.98	34.8 - 47.1	P 40.98	1.78	44.5 - 60.2	P 52.35	2.55	23.0 - 31.1	P 27.03	1.3	13
HemosIL APTT-SP	5	32.3 - 43.7	P 38.03	1.32	26.8 - 36.3	P 31.53	2.59	36.2 - 49.0	P 42.58	2.92	39.3 - 53.2	P 46.23	1.83	24.1 - 32.6	P 28.35	1.62	24
HemosIL SynthASii w/CaCl2	6	34.3 - 46.5	P 40.41	1.73	34.2 - 46.3	P 40.26	3.19	30.2 - 40.8	P 35.5	1.12	43.1 - 58.3	P 50.71	2.04	24.0 - 32.5	P 28.26	1.13	21
Siemens FS, Actin Activated, FSL Activated	7	31.7 - 42.9	P 37.29	1.67	21.5 - 29.1	P 25.29	0.79	24.9 - 33.6	P 29.24	1.42	38.4 - 51.9	P 45.16	2.64	20.1 - 27.2	P 23.61	1.13	80
<b>Initial Grouping by Sensitivity or Principle</b>																	
Very high sensitivity reagents	8	35.1 - 47.5	P 41.33	1.13	27.3 - 36.9	P 32.12	1.98	34.8 - 47.1	P 40.98	1.78	44.5 - 60.2	P 52.35	2.55	23.0 - 31.1	P 27.03	1.3	13
High sensitivity reagents	9	32.4 - 43.8	P 38.13	1.38	26.7 - 36.1	P 31.39	2.63	36.1 - 48.9	P 42.5	2.88	39.5 - 53.5	P 46.51	2.26	24.1 - 32.6	P 28.34	1.59	25
All other methods	10	34.3 - 46.5	P 40.41	1.73	34.2 - 46.3	P 40.26	3.19	30.2 - 40.8	P 35.5	1.12	43.1 - 58.3	P 50.71	2.04	24.0 - 32.5	P 28.26	1.13	21
Moderate sensitivity reagents	11	31.7 - 42.9	P 37.29	1.67	21.5 - 29.1	P 25.29	0.79	24.9 - 33.6	P 29.24	1.42	38.4 - 51.9	P 45.16	2.64	20.1 - 27.2	P 23.61	1.13	80
<b>Total Population</b>																	
Whole Population	12	32.5 - 44.0	P 38.25	2.15	24.9 - 33.6	P 29.26	5.69	28.7 - 38.9	P 33.81	5.82	39.8 - 53.9	P 46.83	3.54	21.7 - 29.4	P 25.53	2.53	141

**Fibrinogen****Initial Grouping by Sensitivity or Principle**

Siemens series	1	193 - 289	P 241.0	20.8	326 - 490	P 408.1	44.6	102 - 154	P 128.1	7.8	194 - 291	P 242.7	18.7	218 - 327	P 272.4	15.3	11
<b>Total Population</b>																	
Whole Population	2	203 - 304	P 253.4	31.7	321 - 482	P 401.6	60.8	108 - 161	P 134.4	12.8	215 - 323	P 268.9	50.0	224 - 336	P 279.8	21.7	17