



# AMERICAN ASSOCIATION OF BIOANALYSTS PROFICIENCY TESTING

11931 Wickchester Ln., Ste 200 - Houston, TX 77043  
800-234-5315 - 281-436-5357 - Fax 713-781-5008

**PARTICIPANT STATISTICS**

**FIRST QUADRIMESTER 2022**

## VIRAL ANTIGEN SCREEN, Waived

Method	Specimen 1			Specimen 2		
	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B
BD Veritor	8	1			9	
Binax NOW Flu A&B, wv	3				3	
Consult Diagnostics	13				12	
ID Now/Alere i	28			2	26	
Quidel Quick Vue Infl A+B	13				13	
Quidel Sofia	30				29	1
Sekisui OSOM A&B	5				5	
<b>Total Population</b>	<b>109</b>	<b>1</b>	<b>0</b>	<b>2</b>	<b>106</b>	<b>1</b>
<b>Flagging</b>		***	***	***		

Method	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
BD Veritor	9		9	
Binax NOW Flu A&B, wv	3		3	
Consult Diagnostics	11	1	11	
ID Now/Alere i	25	3	27	
Quidel Quick Vue Infl A+B	13		13	
Quidel Sofia	30		28	2
Sekisui OSOM A&B	5		5	
<b>Total Population</b>	<b>104</b>	<b>4</b>	<b>104</b>	<b>2</b>
<b>Flagging</b>		***		***

Method	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
Alere Binax Now		9	9	
BD Veritor		4	4	
ID Now/Alere i		5	5	
Quidel QuickVue		4	4	
Quidel Sofia	1	5	6	
<b>Total Population</b>	<b>1</b>	<b>27</b>	<b>28</b>	<b>0</b>
<b>Flagging</b>	***			***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.