



CoaguChek XS, XS Plus, Basic

Name	Line No.	Specimen 1					Specimen 2					No. of Labs
		Range & Type	Mean	SD	CV	Range & Type	Mean	SD	CV			
CoaguChek XS, XS Plus INR 2 spec												
Last Updated: March 5th, 2022 18:13												
Initial Grouping by Reagent												
Roche CoaguChek XS	1	2.34 - 3.51 P	2.927	0.111	3.8%	0.99 - 1.49 P	1.238	0.049	4.0%	49		
Roche CoaguChek XS Plus, Plus/Pro	2	2.36 - 3.55 P	2.956	0.05	1.7%	1.0 - 1.5 P	1.25	0.05	4.0%	11		
Initial Grouping by Sensitivity or Principle												
All methods	3	2.35 - 3.52 P	2.932	0.105	3.6%	0.99 - 1.49 P	1.24	0.049	4.0%	60		
Total Population												
Whole Population	4	2.35 - 3.52 P	2.932	0.105	3.6%	0.99 - 1.49 P	1.24	0.049	4.0%	60		

CoaguChek XS Plus Prothrombin Time 2 spec

Last Updated: March 5th, 2022 21:42

Initial Grouping by Reagent												
Roche CoaguChek XS Plus, Plus/Pro	1	29.4 - 39.8 ? P	34.63	2.51	7.2%	12.2 - 16.5 ? P	14.33	1.66	11.6%	8		
Initial Grouping by Sensitivity or Principle												
All methods	2	28.1 - 38.0 ? P	33.0	5.17	15.7%	12.2 - 16.5 ? P	14.33	1.66	11.6%	10		
Total Population												
Whole Population	3	28.1 - 38.0 P	33.0	5.17	15.7%	12.2 - 16.5 P	14.33	1.66	11.6%	10		