



PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2011

C-REACTIVE PROTEIN

C-REACTIVE PROTEIN	Specimen 1		Specimen 2		No. of Labs
Method	NEG	POS	NEG	POS	
ASI	39	1	1	39	40
Beckman Array nephelometric	1			1	1
Biokit Rhe-jet	3			3	3
Cortez Diagnostics	1			1	1
Fisher Sure-Vue	6			6	6
Immunostics immuno/lex-crp	21		1	20	21
Nerl Diagnostics Accutex	1			1	1
Randox CRP Latex	1			1	1
Seradyn SeraTest Latex	1			1	1
Siemens RapiTex	2			2	2
Stanbio #1130	5			5	5
Teco Diag latex slide test	8			8	8
True-Medix Trutest CRP	1			1	1
Total Population	100	1	2	99	101
Flagging		*****	*****		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.