



**PARTICIPANT STATISTICS**

**SECOND QUADRIMESTER 2011**

**C. difficile Toxin / Antigen Detection**

**C. Difficile Toxin A  
(enterotoxin)**

Method	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Negative	Positive for A or B	Positive	Negative	Positive for A or B	Positive	Negative	Positive for A or B	Positive	Negative	Positive for A or B	Positive	Negative	Positive for A or B	Positive
Fischer Sure-Vue		4	1	5				4	1	5				4	1
Inverness C df Tox A/B I,II		14	1	13	2			14	1	14	1			14	1
Inverness Quik Chek Complet		29	7	36				29	7	36				29	7
Inverness Tox A/B Quik Chek		23	1	24				23	1	24				23	1
Meridian ImmunoCard		33	2	35				33	2	35				33	2
Meridian Pre C diff Tox A/B		7	1	7	1			7	1	8				7	1
Remel Xpect		17	1	18				17	1	18				17	1
TechLab EISA		1		1				1		1				1	
<b>Total Population</b>	<b>0</b>	<b>130</b>	<b>14</b>	<b>141</b>	<b>3</b>	<b>0</b>	<b>0</b>	<b>130</b>	<b>14</b>	<b>143</b>	<b>1</b>	<b>0</b>	<b>0</b>	<b>130</b>	<b>14</b>
<b>Flagging</b>	*****				*****	*****	*****				*****	*****	*****		

**C. Difficile Toxin B  
(cytotoxin)**

Method	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Negative		Positive	Negative		Positive	Negative		Positive	Negative		Positive	Negative		Positive
cell culture			1	1					1	1					1
Diagnostic Hybrids			1	1					1	1					1
Inverness Quik Chek Complet			5	5					5	5					5
<b>Total Population</b>	<b>0</b>		<b>8</b>	<b>9</b>		<b>0</b>	<b>0</b>		<b>8</b>	<b>9</b>		<b>0</b>	<b>0</b>		<b>8</b>
<b>Flagging</b>	*****					*****	*****					*****	*****		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.