C-Reactive Protein, quantitative C-Reactive Protein, quantitative

Name	Line No.	Specimen 1 Range & Type	Mean	SD	Specimen 2 Range & Type	Mean	SD	Specimen 3 Range & Type	Mean	SD	Specimen 4 Range & Type	Mean	SD	Specimen 5 Range & Type	Mean	SD	No. of Labs
C-Reactive Protein, quantitative																	
Initial Grouping by Method																	
Beckman Synchron turbidimet	1	0.1 - 0.9 C	0.47	0.08	2.0 - 3.7 S	2.84	0.29	-			-			-			25
J&J Vitros	2	0.0 - 5.4 S	1.01	1.47	3.8 - 5.1 S	4.45	0.23	-			-			-			26
Roche COBAS Integra	3	0.0 - 0.4 C	0.00	0.00	2.4 - 4.0 S	3.16	0.27	-			-			-			10
Siemens Dimension	4	0.0 - 0.6 C	0.20	0.05	2.9 - 4.2 S	3.55	0.20	-			-			-			32
Initial Grouping by Sensitivity or Principle																	
All wide range methods	5	0.0 - 0.6 S	0.14	0.16	0.0 - 56.2 S	12.75	14.48	-			-			-			20
All direct quantitive meth	6	0.0 - 2.7 S	0.44	0.76	0.0 - 15.0 S	3.91	3.71	-			-			-			128
Total Population																	
Whole Population	7	0.0 - 2.4 S	0.37	0.67	0.0 - 25.8 S	5.06	6.91	-			-			-			169