



**PARTICIPANT STATISTICS**

**SECOND QUADRIMESTER 2011**

**CHLAMYDIA / GC / STREP GROUP B ANTIGEN**

**Chlamydia**

Method	Specimen 1			Specimen 2		Specimen 3		Specimen 4		Specimen 5		Pos Refer for ID
	Neg	Pos	Pos Refer for ID	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
BD ProbeTec ET/Viper		32		31	1	32		32			32	
Gen-Probe APTIMA Combo 2		22	1	21	2	23		22	1		22	1
Gen-Probe Pace 2		6	1	7		7		7			6	1
Inverness Clearview		15		15		15		15			15	
Quidel QuickVue		12		12		12		12			12	
Roche Amplicor	1	8		8	1	9		9		1	8	
<b>Total Population</b>	<b>1</b>	<b>97</b>	<b>2</b>	<b>96</b>	<b>4</b>	<b>100</b>	<b>0</b>	<b>98</b>	<b>2</b>	<b>1</b>	<b>97</b>	<b>2</b>
<b>Flagging</b>	*****			*****		*****		*****		*****		

**Neisseria gonorrhoeae**

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Pos	Pos Refer for ID	Pos	Pos Refer for ID	Neg	Pos	
BD ProbeTec ET/Viper	31		30	1	31		31		31		31
Gen-Probe APTIMA Combo 2	23		23		22	1	22	1	23		23
Gen-Probe Pace 2	7		7		7		7		7		7
Roche Amplicor	9		9		9		9		9		9
<b>Total Population</b>	<b>71</b>	<b>1</b>	<b>71</b>	<b>1</b>	<b>71</b>	<b>1</b>	<b>71</b>	<b>1</b>	<b>72</b>	<b>0</b>	<b>72</b>
<b>Flagging</b>	*****		*****						*****		

**Strep Group B**

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Remel Wellcogen	1			1	1			1	1		1
BD Directigen-latex	3			3	3		3		3		3
Gen-Probe AccuProbe	2			2	2		2		2		2
Remel PathoDX	2			2	2		2		2		2
<b>Total Population</b>	<b>8</b>	<b>0</b>	<b>0</b>	<b>8</b>	<b>8</b>	<b>0</b>	<b>7</b>	<b>1</b>	<b>8</b>	<b>0</b>	<b>8</b>
<b>Flagging</b>	*****		*****				*****		*****		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.