



PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2011

HELICOBACTER PYLORI

H. pylori

Method	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
ABI SureStep		4	4		4
Alfa Instant-View		16	16		16
American Diag Prod Rapid		6	6		6
Beckman Coulter Icon		15	15		15
BioStar Aceava		4	4		4
BioUSA		4	4		4
Cardinal Health SP		12	11	1	12
Consult	1	9	9	1	10
Diag Test Group Clarity		3	3		3
Fisher Sure-Vue IgG		29	26	3	29
Genzyme OSOM	1	2	2	1	3
Germaine		4	4		4
Henry Schein One Step		9	9		9
Instant Technologies iScrn		4	4		4
Inverness Clearview I/II		35	33	2	35
Inverness IgG ELISA		5	5		5
McKesson Medi-Lab		11	7	4	11
Meridian ImmunoCard IgG EIA		6	6		6
Micro Detect IgG		3	3		3
Polymedco Poly stat IgG		15	15		15
Quidel QuickVue gII		83	82	1	83
Select Diagnostics	1	2	3		3
Siemens Immulite 2000		10	10		10
SP Rapid		6	6		6
Stanbio		3	3		3
Total Population	5	318	310	13	322
Flagging	*****			*****	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.