



**PARTICIPANT STATISTICS**

**SECOND QUADRIMESTER 2011**

**RUBELLA**

Rubella IgG Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Abbott Axsym IgG only		9	9		9		9		9		9
ASI Rubella Test	5	27	31	1	30	2	2	30	3	29	32
Beckman Access IgG		12	12		12		12		12		12
Biokit Rbgen/Fisher SureVue		19	19		19		19		19		19
Biokit/SV Select Rubella		3	3		3		3		3		3
bioMerieux Vidas IgG ELFA		6	6		6		6		6		6
BioRad Rubella IgG EIA		2	2		2		2		2		2
Diamedix Microassay IgG		6	6		6		6		6		6
Inverness Rubella-Plus		13	13		13		13		13		13
J&J Vitros		8	8		8		8		8		8
Remel Ruba-Test		3	3		3		3		3		3
Roche Cobas/Elecsys Rub IgG		5	5		5		5	1	4		5
Seradyn Color Slide Rubella		2	2		2		2		2		2
SeraQuest Rubella IgG ELISA		2	2		2		2		2		2
Siemens Immulite 2000, 2500		13	13		13		13		13		13
Siemens Immulite IgG		7	7		7		7		7		7
Siemens Immuno 1/Advia IgG		6	2	3	2	3	6		6		6
True Medix Trutest Rubella		2	2		2		2		2		2
Wampole Impact		4	4		4		4		4		4
Zeus ELISA		2	2		2		2		2		2
<b>Total Population</b>	<b>5</b>	<b>158</b>	<b>158</b>	<b>4</b>	<b>157</b>	<b>5</b>	<b>2</b>	<b>161</b>	<b>4</b>	<b>159</b>	<b>163</b>
<b>Flagging</b>	<b>*****</b>			<b>*****</b>		<b>*****</b>	<b>*****</b>		<b>*****</b>		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.