



PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2011

STREP GROUP A ANTIGEN SCREEN

Strep A Screen Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Abbott Signify		9	9			9		9	9		9
BD Chek		3	3			3		3	3		3
BD Directigen EZ		7	7			7		7	7		7
Beckman Coulter ICON ser		5	5			5		5	5		5
Cardinal Health SP		22	22			22		22	21		22
Cardinal Health SP nonwvd		6	6			6		6	6		6
Consult Twist		4	4			4		4	4		4
Diag Test Group Clarity		3	3			3		3	3		3
Fish Sure-Vue Signature, wv		5	5			5		5	5		5
Fisher Sure-Vue, nonwaived		14	14			14		14	14		14
Fisher Sure-Vue, waived		44	44			44		44	44		44
Genzyme OSOM		24	24			24		24	24		24
Genzyme OSOM Ultra		53	53		1	52	1	52	51	2	53
Genzyme OSOM, waived		21	21			21		21	21		21
Genzyme Rapid Strep A		5	5			5		5	5		5
Henry Schein OneStep		14	14			14		14	14		14
Henry Schein OneStep+, Wv	1	8	8	1		9	1	8	9		9
Immunostics Detector		3	3			3		3	3		3
Immunostics immune/Strep A		3	3			3		3	3		3
Invern/BioStar Acceava		78	78			78		78	77	1	78
Inverness Binax Now		3	3			3		3	3		3
Inverness Clearview		19	19			19		19	19		19
McKesson Twist		11	11			11		11	11		11
McKesson, nonwaived		9	9			9		9	9		9
Meridian Diag ImmunoCard		10	10			10		10	10		10
Polymedco Poly Stat		5	5			5		5	5		5
PSS Select / Consult Diag		20	20			20		20	20		20
Quidel QuickVue +		5	5			5		5	5		5
Quidel QuickVue +	1	66	67		1	66	1	66	66	1	67
Quidel QuickVue Dipstick		41	40	1	1	40		41	40		41
Quidel QuickVue In-Line	7	111	118	1	3	116	6	113	117	1	119
SA Scientific Strep Alert		3	3			3		3	3		3
Select Diagnostics		4	3	1	1	2		4	4		4
Total Population	11	663	670	5	7	667	10	665	667	5	675
Flagging	*****			*****	*****		*****			*****	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.