



PARTICIPANT STATISTICS

VIRAL ANTIGEN SCREEN

SECOND QUADRIMESTER 2016

Influenza Type A

Method	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			
	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	
Alere i	6			6				6			6			6		
BD Veritor	14			14				14			13	1		14		
Binax NOW Flu A&B	61			61			1	60			59	2		61		
Meridian Tru-Flu	6			6				6			6			6		
Quidel Quick Vue Infl A+B	22			22				22			22			22		
Quidel Sofia Infl A+B	22			21	1			22			22			22		
Remel Xpect Flu A&B	23			23				23			23			23		
Sekisui OSOM A&B	16			16				16			16			16		
Total Population	177	0	0	176	1	0	1	176	0	172	5	0	176	1	0	
Flagging		***	***		***	***	***				***			***	***	***

Influenza Type B

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Alere i	5	1	6		5	1		6	4	2
BD Veritor	14		14		14		2	12	13	1
Binax NOW Flu A&B	60		60		59	1	1	59	58	1
Meridian Tru-Flu	6		6		6			6	6	
Quidel Quick Vue Infl A+B	22		22		22			22	21	1
Quidel Sofia Infl A+B	22		21	1	22		2	20	21	1
Remel Xpect Flu A&B	23		23		23			23	23	
Sekisui OSOM A&B	16		16		16			16	16	
Total Population	175	1	175	1	174	2	5	171	168	7
Flagging		***		***		***	***			***

Adenovirus

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
RPS		1	1		1		1		1	
SA Scientific SAS		1	1		1		1		1	
Total Population	0	4	3	1	3	1	4	0	4	0
Flagging	***							***		***

*Due to a lack of participant consensus, Specimens 2 and 3 were not evaluated this event.

Respiratory Syncytial

Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Alere Binax Now		45	45		46		46		46	
BD Veritor		7	7		7		7		7	
BD Veritor, non-waived		4	4		4		4		4	
Fisher Sure-Vue RSV Test		3	3		3		3		3	
Meridian Tru RSV		6	6		6		6		6	
Quidel QuickVue		10	10		10		10		10	
Quidel Sofia		8	8		8		8		8	
Total Population	0	92	92	0	92	1	93	0	93	0
Flagging	***			***		***		***		***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.