



PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2016

HELICOBACTER PYLORI

H. pylori

Method	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
Alere Clearview		20	20		20
Alfa Scientific InstantView	2	21	20	3	23
Beckman Coulter Icon		5	5		5
Cardinal Health SP		14	14		14
Consult Diagnostics	1	7	7	1	8
Fisher Sure-Vue IgG		9	9		9
Henry Schein One Step		7	6	1	7
McKesson Medi-Lab		4	4		4
Micro Detect IgG		5	5		5
Polymedco Poly stat IgG		5	5		5
Quidel QuickVue gII		29	29		29
Siemens Immulite 2000		4	4		4
Total Population	4	156	154	6	160
Flagging	***			***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.