



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. • Brownsville, TX 78520-5596
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2016

C-Reactive Protein, quantitative

Name	Line No.	Specimen 1				Specimen 2				No. of Labs
		Range	Type	Mean	SD	Range	Type	Mean	SD	
C-Reactive Protein, quantitative										
Initial Grouping by Method										
Ortho Vitros	1	0 - 4.6	S	0.82	1.27	0.7 - 3.3	S	2.03	0.44	13
Siemens CRP Extended RCRP	2	0 - 0.4	C	0.0	0.0	1.0 - 1.9	C	1.45	0.05	4
Roche COBAS Integra	3	0 - 0.4	C	0.02	0.04	0.9 - 1.7	C	1.34	0.05	5
Siemens Dimension	4	0 - 0.6	C	0.18	0.06	1.1 - 2.0	S	1.52	0.15	32
Roche Tina-Quant / CRPL3	5	0 - 3.2	S	0.39	0.93	0 - 12.3	S	2.32	3.32	11
Abbott Aeroset/Architect	6	0 - 4.8	S	0.48	1.43	0 - 2.7	S	1.12	0.54	11
Beckman Olympus turb	7	0 - 1.2	S	0.25	0.32	0.5 - 2.4	S	1.41	0.32	21
Beckman Synchron turbidimet	8	0.1 - 0.9	C	0.54	0.08	0.7 - 1.5	S	1.14	0.14	6
Initial Grouping by Sensitivity or Principle										
All direct quantitative meth	9	0 - 2.8	S	0.35	0.82	0 - 5.2	S	1.57	1.2	88
All wide range methods	10	0 - 1.1	S	0.19	0.29	0.6 - 2.2	S	1.42	0.27	32
Total Population										
Whole Population	11	0 - 14.5	S	0.79	4.57	0 - 30.7	S	2.95	9.25	124