RETICULOCYTE

Name	Line No.	Specimen 1 Range & Type	Specimen 1 Mean	Specimen 1 SD	Specimen 2 Range & Type	Specimen 2 Mean	Specimen 2 SD	Specimen 3 Range & Type	Specimen 3 Mean	Specimen 3 SD	Specimen 4 Range & Type	Specimen 4 Mean	Specimen 4 SD	Specimen 5 Range & Type	Specimen 5 Mean	Specimen 5 SD	No. of Labs
Reticulocyte Count, automated																	
Initial Grouping by Reagent																	
Sysmex X series	1	6.0 - 11.2	P 8.63	1.54	2.9 - 5.4	P 4.18	0.62	-			-			-			33
Beckman Coulter LH750	2	4.9 - 9.1	P 7.02	1.1	2.5 - 4.7	P 3.62	0.22	-			-			-			9
Coulter DxH 800/600	3	4.8 - 9.0	P 6.91	0.38	2.4 - 4.4	P 3.4	0.13	-			-			-			9
Initial Grouping by Sensitivity or Principle																	
Sysmex, Cell-Dyn 3000 ser	4	6.0 - 11.1	P 8.53	1.62	2.9 - 5.4	P 4.13	0.67	-			-			-			34
Coulter	5	5.0 - 9.2	P 7.08	0.86	2.5 - 4.7	P 3.62	0.35	-			-			-			21
Total Population																	
Whole Population	6	5.6 - 10.4	P 8.0	1.52	2.8 - 5.2	P 4.03	0.92	-			-			-			58
Reticulocyte Count, manual																	
Initial Grouping by Reagent																	
Manual	1	4.6 - 9.6	C 7.13	1.47	1.1 - 6.1	C 3.64	1.22	-			-			-			27
Manual - Miller Ocular	2	2.8 - 7.8	C 5.33	0.88	0.1 - 5.2	C 2.65	0.77	-			-			-			4
Initial Grouping by Sensitivity or Principle																	
Manual	3	4.6 - 9.6	C 7.13	1.47	1.1 - 6.1	C 3.64	1.22	-			-			-			27
Manual - Miller Ocular	4	2.8 - 7.8	C 5.33	0.88	0.1 - 5.2	C 2.65	0.77	-			-			-			4
Total Population																	
Whole Population	5	4.6 - 9.6	C 7.1	1.88	1.0 - 6.0	C 3.51	1.21	-			-			-			31
Reticulocyte manual, Digital Slide																	
Total Population																	
Whole Population	1	0 - 55.8	P 1.8	0.0	0 - 186.0	P 6.0	0.0	-			-			-			1