



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. w Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**ROTAVIRUS ANTIGEN SCREEN**

**SECOND QUADRIMESTER 2017**

Rotavirus Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Fisher Sure-Vue Rota Test		3	3		3		3			3	3
Meridian ImmunoCard Stat		7	7		7		7			6	7
Remel Xpect Rota Test		1	1		1		1		1		1
SA Scientific SAS		2	2		2		2			2	2
<b>Total Population</b>	<b>0</b>	<b>13</b>	<b>13</b>	<b>0</b>	<b>13</b>	<b>0</b>	<b>13</b>	<b>0</b>	<b>1</b>	<b>11</b>	<b>13</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>		<b>***</b>		<b>***</b>	<b>***</b>		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.