



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. w Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**SECOND QUADRIMESTER 2018**

**GENITAL CULTURE**

<b>GC Culture</b> Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Labs	Flagging	Labs	Flagging	Labs	Flagging	Labs	Flagging	Labs	Flagging
Negative for Gram negative diplococci	0		0	***	0	***	0		0	***
Positive for Gram negative diplococci	0	***	1	***	1	***	1	***	1	
Negative for Neisseria gonorrhoeae by direct antigen screen	0		0	***	0	***	0		0	***
Positive for Neisseria gonorrhoeae by direct antigen screen	0	***	0		0		0	***	0	
Negative for Neisseria Species	0		0	***	0	***	0		0	***
Positive for Neisseria species	0	***	0		0		0	***	0	
Negative for Neisseria gonorrhoeae by culture	8		8	***	8	***	8		0	***
Positive for Neisseria gonorrhoeae by culture	0	***	0		0		0	***	6	
Positive for Neisseria gonorrhoeae (beta-lactamase positive) by culture	0	***	0		0		0	***	2	
No aerobic growth	2		1	***	1	***	1		1	***
<b>Total Population</b>	<b>10</b>		<b>10</b>		<b>10</b>		<b>10</b>		<b>10</b>	

<b>Group B Strep</b> Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg
<b>Total Population</b>	<b>2</b>	<b>0</b>	<b>0</b>	<b>2</b>	<b>0</b>	<b>2</b>	<b>0</b>	<b>2</b>	<b>0</b>	<b>2</b>
<b>Flagging</b>		<b>***</b>	<b>***</b>		<b>***</b>		<b>***</b>		<b>***</b>	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.