



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**
205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2018

CHLAMYDIA / GC / STREP GROUP B ANTIGEN

Chlamydia	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD Max		4		3	1		3	1			4		3	1	
BD ProbeTec ET/Viper	2	18		19	1		20				20		19	1	
Cepheid		19		17	2		18				19		16	3	
Gen-Probe APTIMA Combo		13		10	3		12	1			13		12	1	
Quidel QuickVue		5		5			5				5		5		
Total Population	2	61	0	56	7	0	60	2	0	0	63	0	57	6	0
Flagging	***		***		***	***		***	***	***		***		***	***

Neisseria gonorrhoeae	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD Max	4			4				4		3	1				4
BD ProbeTec ET/Viper	20			20			1	19		20			1	19	
Cepheid	18	1		18	1			18		16	3			19	
Gen-Probe APTIMA Combo	13			13				13		12	1			13	
Total Population	57	1	0	57	1	0	1	56	0	53	5	0	1	57	0
Flagging		***	***		***	***	***				***		***		

Strep Group B	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
BD Max		3	3		2	1		3	3	
Cepheid		4	4		4			4	4	
Total Population	0	12	12	0	11	1	1	11	12	0
Flagging	***			***		***	***			***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.