



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2018

HELICOBACTER PYLORI

<i>H. pylori</i>	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
Alere Clearview		14	14		14
Alere ELISA		3	3		3
Alfa Scientific InstantView		18	18		18
BTNX Rapid Response, wv	1	6	6	1	7
Cardinal Health SP		7	7		7
Consult Diagnostics		6	6		6
Fisher Sure-Vue IgG		7	7		7
Henry Schein One Step		8	8		8
Micro Detect IgG		4	4		4
Quidel QuickVue gII		23	23		23
Siemens Immulite 2000		3	3		3
Total Population	1	122	122	1	123
Flagging	***			***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.