



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. w Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2018

STREP GROUP A ANTIGEN SCREEN

Step A Screen Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Binax Now		3		3	3		3		3		3
Alere BioSta Acceava wvd		15		15	15		15		15		15
Alere i		7		7	7		7		7		7
BD Veritor		6		6	6		6		6		6
BTNX Rapid Response		3		3	3		3		3		3
Cardinal Health SP		5		5	5		5		5		5
Consult Dipstick, waived		11	1	10	11		11		11		11
Fisher Sure-Vue, waived		10		10	10		10		10		10
Henry Schein OneStep		4		4	4		4		4		4
Henry Schein OneStep+, Wv		3		3	3		3		3		3
McKesson, waived		6		6	6		6		6		6
Meridian Diag ImmunoCard		3		3	3		3		3		3
Quidel QuickVue +		18		18	18		18		18		18
Quidel QuickVue Dipstick		13		13	13		13		13		13
Quidel QuickVue In-Line	1	38		39	38	1	39		39		39
Quidel Sofia		11		11	11		11		11		11
Sekisui OSOM		9		9	9		9		9		9
Sekisui OSOM Ultra		15		15	15		15		15		15
Sekisui OSOM, waived		20		20	20		20		20		20
Sekisui Rapid Strep A		3		3	3		3		3		3
Total Population	2	217	2	217	218	1	219	0	219	0	219
Flagging	***		***			***		***		***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.