



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2018

Fetal RBCs

Fetal RBCs Method	Specimen 1		Specimen 2		Specimen 3	
	NEG	POS	NEG	POS	NEG	POS
Immucor FMH		5	1	4	2	3
Total Population	0	5	1	4	2	3
Flagging	***		***			

*Due to a lack of participant consensus, Specimen 3 was not evaluated this event. The intended result was Pos.

Fetal RBCs, Quant

	Line No.	Specimen 1			Specimen 2			Specimen 3			No. of Labs			
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD				
Total Population														
Whole Population	1	0-0.33	C	0.3	0	4.53-5.13	C	4.83	0	0-0.33	C	0.3	0	1

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.