



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2019

C. difficile Toxin / Antigen Detection

**C. Difficile Toxin A
(enterotoxin)**

Method	Specimen 6			Specimen 7			Specimen 8			Specimen 9			Specimen 10			
	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B	
Alere C df Tox A/B I,II	2			2				2			2				2	
Alere Quik Chek Complet	26			26				8	18	26					8	18
Alere Tox A/B Quik Chek	7			7				2	5	7					2	5
Meridian Illumigene	4			4				2	2	4					2	2
Meridian ImmunoCard	10			10				1	9	10					1	9
Quidel AmpliVue	2			2				1	1	2			1		1	
Remel Xpect	2			2					2	2						2
Total Population	57	0	0	57	0	0	0	16	41	57	0	0	1	16	40	
Flagging		***	***		***	***	***				***	***	***			

**C. Difficile Toxin B
(cytotoxin)**

Method	Specimen 6		Specimen 7		Specimen 8		Specimen 9		Specimen 10	
	Negative	Positive	Negative	Positive	Negative	Positive	Negative	Positive	Negative	Positive
Alere Quik Chek Complet	4		4			4	4			4
Cepheid	4		4			4	4			4
Total Population	9	0	9	0	0	9	9	0	0	9
Flagging		***		***	***			***	***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.