



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**SECOND QUADRIMESTER 2019**

**C-Reactive Protein, qualitative**

C-REACTIVE PROTEIN Method	Specimen 6		Specimen 7		No. of Labs
	NEG	POS	NEG	POS	
ASI		21	21		21
Immunostics immuno/lex-crp		6	6		6
Teco Diag latex slide test		4	4		4
<b>Total Population</b>	<b>0</b>	<b>39</b>	<b>39</b>	<b>0</b>	<b>39</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.