



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2019

HELICOBACTER PYLORI

<i>H. pylori</i>	Specimen 6		Specimen 7		No. of Labs
	Neg	Pos	Neg	Pos	
Alere Clearview		12	12		12
Alere ELISA		3	3		3
Alfa Scientific InstantView		19	19		19
BTNX Rapid Response, wv		6	6		6
Cardinal Health SP		5	5		5
Consult Diagnostics		8	8		8
Fisher Sure-Vue IgG		7	7		7
Henry Schein One Step		8	8		8
Micro Detect IgG		3	3		3
Quidel QuickVue gII		18	18		18
Total Population	0	111	111	0	111
Flagging	***			***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.