



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2019

HIV-1 Waived/Rapid

anti-HIV-1 Waived

Method	Specimen 6			Specimen 7		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere Determine	13	1			14	
Chembio HIV 1/2 STAT-PAK	4				4	
Total Population	20	1	1	0	22	0
Flagging		***	***	***		***

p24 Antigen

Method	Specimen 6			Specimen 7		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere Determine		14		14		
Total Population	1	14		14	1	
Flagging	***				***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.