



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. w Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**SECOND QUADRIMESTER 2019**

**RUBELLA**

Rubella IgG Method	Specimen 6		Specimen 7		Specimen 8		Specimen 9		Specimen 10		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Rubella-Plus		7	7		1	6	7		1	6	7
ASI Rubella Test		15	15			15	15			15	15
Beckman Access IgG		10	10			10	10			10	10
Bio-Rad Bioplex 2000, 2200		5	5			5	5			5	5
Fisher Sure-Vue		8	8		1	7	7	1		8	8
Roche Elecsys		5	5			5	5			5	5
Siemens Immulite 2000, 2500		6	6		1	5	5	1		6	6
Siemens Immuno 1/Advia IgG		3	3			3	2	1		3	3
<b>Total Population</b>	<b>0</b>	<b>83</b>	<b>83</b>	<b>0</b>	<b>3</b>	<b>80</b>	<b>80</b>	<b>3</b>	<b>1</b>	<b>82</b>	<b>83</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>	<b>***</b>			<b>***</b>	<b>***</b>		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.