



**PROFICIENCY TESTING SERVICE  
AMERICAN ASSOCIATION OF BIOANALYSTS**  
205 West Levee St. ♦ Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**SECOND QUADRIMESTER 2019**

**Strep A Screen, waived**

Method	Specimen 6		Specimen 7		No. of Labs
	Neg	Pos	Neg	Pos	
Alere i	16			17	16
Alere/BioSta Acceava, Wv	12			12	12
BD Veritor	11			11	11
Cardinal Health SP	3			3	3
Fisher Sure-Vue, waived	3			3	3
Henry Schein OneStep+, Wv	6			6	6
Mckesson, Consult waived	12			12	12
Meridian Diag ImmunoCard	5			5	5
NDC Pro Advantage	3			3	3
PSS Select / Consult Diag	7			7	7
Quidel QuickVue Dipstick	9			9	9
Quidel QuickVue In-Line	31	1	4	28	32
Quidel Sofia	18			18	18
Sekisui OSOM, waived	29			29	29
<b>Total Population</b>	<b>173</b>	<b>1</b>	<b>4</b>	<b>171</b>	<b>174</b>
<b>Flagging</b>		***	***		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.