



**AMERICAN ASSOCIATION OF BIOANALYSTS
PROFICIENCY TESTING**

11931 Wickchester Ln., Ste 200 - Houston, TX 77043
800-234-5315 - 281-436-5357 - Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND EVENT 2022

C. difficile Toxin / Antigen Detection

**C. Difficile Toxin A
(enterotoxin)**

Method	Specimen 6			Specimen 7			Specimen 8			Specimen 9			Specimen 10		
	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B	Negative	Positive	Positive for A or B
Alere Quik Chek Complet	16			16				6	10		6	10		6	10
Alere Tox A/B Quik Chek	3			3				1	2		1	2		1	2
Cepheid	3			3				1	2		1	2		1	2
Meridian ImmunoCard	6			6				1	6		1	6		1	6
Meridian Pre C diff Tox A/B	2			2					2			2			2
TechLab EISA	3			3				2	1		2	1		2	1
Total Population	36	0	0	36	0	0	0	12	25	0	12	25	1	11	25
Flagging		***	***		***	***	***			***			***		

**C. Difficile Toxin B
(cytotoxin)**

Method	Specimen 6		Specimen 7		Specimen 8		Specimen 9		Specimen 10	
	Negative	Positive	Negative	Positive	Negative	Positive	Negative	Positive	Negative	Positive
Alere Quik Chek Complet	5		5			5		5		5
Cepheid	2		2			2		2		2
Total Population	10	0	10	0	0	9	0	9	0	9
Flagging		***		***	***		***		***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.