



PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2011

C-REACTIVE PROTEIN

C-REACTIVE PROTEIN Method	Specimen 1		Specimen 2		No. of Labs
	NEG	POS	NEG	POS	
ASI	40			40	40
Beckman Array nephelometric	1			1	1
Biokit Rhe-jet	2			2	2
Cortez Diagnostics	1			1	1
Fisher Sure-View	8			8	8
Nerl Diagnostics Accutex	1			1	1
Immunostics immuno/lex-crp	19			19	19
Randox CRP Latex	1			1	1
Seradyn SeraTest Latex	2			2	2
Siemens RapiTex	2			2	2
Stanbio #1130	5		1	4	5
Teco Diag latex slide test	9			9	9
True-Medix Trutest CRP	1			1	1
Total Population	100	0	1	99	100
Flagging		*****			

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.