C-Reactive Protein, quantitative C-Reactive Protein, quantitative

Name	Line No.	Specimen 1 Range & Type		Mean	SD	Specimen 2 Range & Type		Mean	SD	Specimen 3 Range & Type	Mean	SD	Specimen 4 Range & Type	Mean	SD	Specimen 5 Range & Type	Mean	SD	No. of Labs
C-Reactive Protein, quantitative																			
Initial Grouping by Method																			
Beckman Synchron turbidimet	1	0.1 - 0.9	C	0.47	0.08	0.7 - 1.6	S	1.17	0.15	-			-			-			25
J&J Vitros	2	0.1 - 0.9	C	0.50	0.13	1.7 - 3.3	S	2.47	0.26	-			-			-			25
Siemens Dimension	3	0.0 - 0.6	C	0.19	0.03	0.2 - 2.1	S	1.18	0.32	-			-			-			32
Initial Grouping by Sensitivity or Principle																			
All wide range methods	4	0.0 - 0.7	S	0.12	0.19	0.0 - 27.8	S	6.36	7.14	-			-			-			18
All direct quantitive meth	5	0.0 - 1.9	S	0.36	0.51	0.0 - 6.5	S	1.68	1.62	-			-			-			129
Total Population																			
Whole Population	6	0.0 - 1.8	S	0.32	0.48	0.0 - 12.7	S	2.35	3.46	-			-			-			168