



PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2011

HELICOBACTER PYLORI

H. pylori

Method	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
ABI SureStep	3			3	3
Alfa Instant-View	13			13	13
American Diag Prod Rapid	7			7	7
Beckman Coulter Icon	15		1	14	15
BioStar Acceava	4			4	4
BioUSA	4			4	4
Cardinal Health SP	14	1	2	13	15
Consult	8			8	8
Diag Test Group Clarity	4			4	4
Fisher Sure-Vue IgG	27		1	26	27
Genzyme OSOM	4			4	4
Germaine	5			5	5
Henry Schein One Step	8			8	8
Instant Technologies iScrn	4			4	4
Inverness Clearview I/II	38		1	37	38
Inverness IgG ELISA	5			5	5
McKesson Medi-Lab	11			11	11
Meridian ImmunoCard IgG EIA	6		1	5	6
Micro Detect IgG	3			3	3
Polymedco Poly stat IgG	14		1	13	14
Quidel QuickVue gII	81	1		82	82
Select Diagnostics	3			3	3
Siemens Immulite 2000	10			10	10
SP Rapid	5	1		6	6
Stanbio	3			3	3
Total Population	319	3	7	315	322
Flagging		*****	*****		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.