



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

CHLAMYDIA / GC / STREP GROUP B ANTIGEN

THIRD QUADRIMESTER 2016

Chlamydia	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD ProbeTec ET/Viper	12	17		29			29			1	28		29		
Cepheid		17		16			17				17		17		
Gen-Probe APTIMA Combo 2	1	13		14			14				14		14		
Quidel QuickVue	9			9			9				9		9		
Total Population	23	50	0	72	0	0	73	0	0	1	72	0	73	0	0
Flagging					***	***		***	***	***				***	***

*Due to a lack of participant consensus, Specimen 1 was not evaluated.

Neisseria gonorrhoeae	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD ProbeTec ET/Viper	29			29			4	25		29			4	25	
Cepheid	17			17				17		17				17	
Gen-Probe APTIMA Combo 2	14			14			12	2		14			12	2	
Total Population	64	0	0	64	0	0	17	47	0	64	0	0	17	47	0
Flagging		***	***		***	***					***	***			

*Due to a lack of participant consensus, Specimens 3 and 5 were not evaluated.

Strep Group B	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Total Population	1	7	1	7	8	0	8	0	8	0
Flagging	***		***			***		***		***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.