



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**THIRD QUADRIMESTER 2018**

**CHLAMYDIA / GC / STREP GROUP B ANTIGEN**

Chlamydia	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD Max		4			4		4			4			4		
BD ProbeTec ET/Viper	1	17		1	17		17	1		17	1		17	1	
Cepheid		20			20		19			19	1		20		
Gen-Probe APTIMA		13			13		13			12	1		12	1	
Quidel QuickVue		5			5		5			5			5		
<b>Total Population</b>	<b>1</b>	<b>61</b>	<b>0</b>	<b>1</b>	<b>61</b>	<b>0</b>	<b>60</b>	<b>1</b>	<b>0</b>	<b>59</b>	<b>3</b>	<b>0</b>	<b>60</b>	<b>2</b>	<b>0</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>				<b>***</b>	<b>***</b>		<b>***</b>	<b>***</b>		<b>***</b>	

Neisseria gonorrhoeae	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer	Neg	Pos	Positive for Chlamydia or GC, would refer
BD Max	4				4		4			4				4	
BD ProbeTec ET/Viper	18			1	17		18			18			1	17	
Cepheid	19	1			20		18	1		18	2			20	
Gen-Probe APTIMA	13			1	12		13			13				13	
<b>Total Population</b>	<b>56</b>	<b>1</b>	<b>0</b>	<b>2</b>	<b>55</b>	<b>0</b>	<b>55</b>	<b>1</b>	<b>0</b>	<b>55</b>	<b>2</b>	<b>0</b>	<b>1</b>	<b>56</b>	<b>0</b>
<b>Flagging</b>		<b>***</b>		<b>***</b>				<b>***</b>	<b>***</b>		<b>***</b>	<b>***</b>	<b>***</b>		

Strep Group B	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
BD Max	3		3		3		3			3
Cepheid	4		4		4		4			4
<b>Total Population</b>	<b>12</b>	<b>0</b>	<b>12</b>	<b>0</b>	<b>12</b>	<b>0</b>	<b>12</b>	<b>0</b>	<b>0</b>	<b>12</b>
<b>Flagging</b>		<b>***</b>		<b>***</b>		<b>***</b>		<b>***</b>	<b>***</b>	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.