



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2018

HIV-1 Waived/Rapid

anti-HIV-1 Waived	Specimen 1			Specimen 2		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere Determine		19		19		
Chembio HIV 1/2 STAT-PAK		4		4		
OraSure OraQuick ADVANCE, Wv		6		6		
OraSure OraQuick Rapid		3		2		1
Total Population	0	35	0	34	0	1
Flagging	***		***		***	***

p24 Antigen	Specimen 1			Specimen 2		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere Determine	19				19	
Total Population	19	1	0	1	19	0
Flagging		***	***	***		***

* Alere Determine users appear to correctly have identified the positive p24 Antigen, but have largely called the screen negative. Methods that report antigen and anti-HIV separately should have reported Negative for Specimen 2.

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.