



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
205 West Levee St. ♦ Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**THIRD QUADRIMESTER 2018**

**Infectious Mononucleosis Waived**

<b>Infectious Mono</b>	<b>Specimen 1</b>		<b>Specimen 2</b>	
	<b>Neg</b>	<b>Pos</b>	<b>Neg</b>	<b>Pos</b>
Alere Acceava	1	11	11	1
Alere Clearview		15	15	
Biokit/Fisher Sure-		3	3	
Consult Diagnostics		10	10	
Henry Schein One		3	3	
Sekisui OSOM Mono		13	13	
<b>Total Population</b>	<b>1</b>	<b>67</b>	<b>67</b>	<b>1</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees.  
Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.