



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
205 West Levee St. ♦ Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**THIRD QUADRIMESTER 2018**

**Rapid Urease**

Method	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
CLotest		8	8		8
<b>Total Population</b>	<b>0</b>	<b>10</b>	<b>10</b>	<b>0</b>	<b>10</b>
<b>Flagging</b>	<b>***</b>			<b>***</b>	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.