



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. w Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2018

RUBELLA

Rubella IgG Method	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Rubella-Plus	7			7	7		7		7		7
ASI Rubella Test	18			18		18	17	1	17	1	18
Beckman Access IgG	10	1	1	10		11	10	1	11		11
Biokit Rbgen/Fisher SureVue	5			5		5	5		5		5
Biokit Rubajet	3			3		3	3		3		3
bioMerieux Vidas IgG ELFA	3			3		3	3		2	1	3
Bio-Rad Bioplex 2000, 2200	4			4		4	4		4		4
Fisher Sure-Vue	6			6		6	6		6		6
Roche Elecsys	5			5		5	5		5		5
Siemens Immulite 2000, 2500	6			6		6	6		6		6
Siemens Immuno 1/Advia IgG	4			4		4	4		4		4
Total Population	85	1	1	85	0	86	84	2	84	2	86
Flagging		***	***		***			***		***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.