



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**THIRD QUADRIMESTER 2018**

**Strep A Screen, waived**

Method	Specimen 1		Specimen 2		No. of Labs
	Neg	Pos	Neg	Pos	
Alere i	28			27	28
Alere/BioSta Acceava, Wv	11			11	11
BD Veritor	10			10	10
Cardinal Health SP	3			3	3
Henry Schein OneStep+, Wv	8			8	8
Mckesson, Consult waived	12			12	12
Meridian Diag ImmunoCard	11			11	11
PSS Select / Consult Diag	5			5	5
Quidel QuickVue Dipstick	11			11	11
Quidel QuickVue In-Line	32		1	31	32
Quidel Sofia	25			25	25
Sekisui OSOM, waived	31			31	31
<b>Total Population</b>	<b>202</b>	<b>0</b>	<b>2</b>	<b>199</b>	<b>202</b>
<b>Flagging</b>		***	***		

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.