



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. ♦ Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**ToRCH**

**THIRD QUADRIMESTER 2018**

Cytomegalovirus	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
<b>Total Population</b>	0	7	7	0	3	4
<b>Flagging</b>	***			***		

\*Due to a lack of participant consensus, Specimen 3 was not evaluated this event. The intended result was Positive.

Herpes Type II or I/II, IgG	Specimen 1			Specimen 2			Specimen 3		
	Neg	Pos Type II	Pos for I or II	Neg	Pos Type II	Pos for I or II	Neg	Pos Type II	Pos for I or II
DiaSorin LIAISON		3			3			3	
Focus Diagnostics		4			4			4	
<b>Total Population</b>	0	13	1	2	11	1	2	11	1
<b>Flagging</b>	***			***			***		

Herpes Type I, IgG or non-spec	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
DiaSorin LIAISON		3		3	3	
Focus Diagnostics		4		4	4	
<b>Total Population</b>	1	13	1	13	10	4
<b>Flagging</b>	***		***			

\*Due to a lack of participant consensus, Specimen 3 was not evaluated this event. The intended result was Negative.

Herpes Type II or I/II, IgM	Specimen 1			Specimen 2			Specimen 3		
	Neg	Pos	Pos I/II	Neg	Pos	Pos I/II	Neg	Pos	Pos I/II
<b>Total Population</b>	6	1	0	7	0	0	7	0	0
<b>Flagging</b>		***	***		***	***		***	***

Herpes Type I, IgM	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
<b>Total Population</b>	3	0	3	0	3	0
<b>Flagging</b>		***		***		***

Rubella IgM	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
DiaSorin LIAISON		11	11		11	
<b>Total Population</b>	2	19	21	0	21	0
<b>Flagging</b>	***			***		***

Toxoplasma	Specimen 1		Specimen 2		Specimen 3	
	Neg	Pos	Neg	Pos	Neg	Pos
<b>Total Population</b>	6	0	6	0	5	1
<b>Flagging</b>		***		***		***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.