



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. w Brownsville, TX 78520-5596  
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**VIRAL ANTIGEN SCREEN**

**THIRD QUADRIMESTER 2018**

Influenza Type A	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B
Alere	4				4		4			4			4		
Alere i	14				14		11	3		13	1		13		1
BD Veritor	31				31		31			31			31		
Binax NOW Flu A&B	6				6		6			5	1		6		
Henry Schein One Step	3				3		3			2	1		3		
Meridian ImmunoCard STAT	4				4		4			4			4		
OraSure	17				17		17			17			17		
Quidel Quick Vue Infl A+B	4				4		4			4			4		
Quidel Sofia Infl A+B	23				23		21	2		22	1		23		
Remel Xpect Flu A&B	4				4		4			4			4		
Sekisui OSOM A&B	3				3		3			3			3		
<b>Total Population</b>	<b>125</b>	<b>1</b>	<b>0</b>	<b>0</b>	<b>125</b>	<b>1</b>	<b>121</b>	<b>5</b>	<b>0</b>	<b>121</b>	<b>4</b>	<b>1</b>	<b>125</b>	<b>1</b>	<b>0</b>
<b>Flagging</b>		***	***	***				***	***		***			***	***

Influenza Type B	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
Method	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Alere i	12	2	11	3	13	1		14	11	3
BD Veritor	31		31		31		31	31		
OraSure	17		17		17			17		
Quidel Sofia Infl A+B	24		24		24		24	21		3
<b>Total Population</b>	<b>124</b>	<b>2</b>	<b>122</b>	<b>4</b>	<b>124</b>	<b>2</b>	<b>0</b>	<b>126</b>	<b>120</b>	<b>6</b>
<b>Flagging</b>		***		***		***	***			***

Adenovirus	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
Method	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
<b>Total Population</b>	<b>0</b>	<b>2</b>	<b>2</b>	<b>0</b>	<b>2</b>	<b>0</b>	<b>2</b>	<b>0</b>	<b>2</b>	<b>0</b>
<b>Flagging</b>	***			***		***		***		***

Respiratory Syncytial Virus	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
Method	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos
Alere Binax Now		21	20	1	21		20	1	20	1
Alere i		6	6		6		5	1	6	
BD Veritor		3	3		3		3		3	
BD Veritor, non-waived		3	3		3		3		3	
Binax NOW RSV		7	7		7		7		7	
Meridian Tru RSV		4	4		4		4		4	
Quidel QuickVue		11	11		11		11		11	
Quidel Sofia		10	10		10		10		10	
<b>Total Population</b>	<b>0</b>	<b>74</b>	<b>73</b>	<b>1</b>	<b>74</b>	<b>0</b>	<b>72</b>	<b>2</b>	<b>73</b>	<b>1</b>
<b>Flagging</b>	***			***		***		***		***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.