



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2018

VIRAL ANTIGEN SCREEN, Waived

Influenza Type A Method	Specimen 1			Specimen 2		
	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B
Alere	14			1	13	
Alere i	26			1	25	
BD Veritor	49	3		1	50	1
Binax NOW Flu A&B, wv	10				10	
Consult Diagnostics	9				9	
Henry Schein One Step +	4				4	
OraSure	6				6	
Quidel Quick Vue Infl A+B	9	1			10	
Quidel Sofia	29				28	1
Sekisui OSOM A&B	10	1		2	9	
Total Population	173	5	1	6	170	3
Flagging		***	***	***		

Influenza Type B Method	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
Alere	14		13	1
Alere I	26		24	2
BD Veritor	52		49	3
Binax NOW Flu A&B, wv	10		10	
Consult Diagnostics	9		9	
Henry Schein One Step +	4		4	
OraSure	7		7	
Quidel Quick Vue Infl A+B	11		11	
Quidel Sofia	28		27	1
Sekisui OSOM A&B	11	1	12	
Total Population	175	1	169	7
Flagging		***		***

Respiratory Syncytial Virus Method	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
Alere Binax Now		15	15	
Alere i		6	6	
BD Veritor		7	6	1
Quidel QuickVue		10	10	
Quidel Sofia		4	4	
Total Population	0	44	43	1
Flagging	***			***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.