



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2018

Fetal RBCs

Fetal RBCs Method	Specimen 1		Specimen 2		Specimen 3	
	NEG	POS	NEG	POS	NEG	POS
Immucor FMH	2	3	5			5
Total Population	2	3	5	0	0	5
Flagging				***	***	

* Due to a lack of participant consensus, Specimen 1 was not evaluated. The intended result was Positive.

Fetal RBCs, Quant

	Line No.	Specimen 1			Specimen 2			Specimen 3			No. of Labs			
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD				
Total Population														
Whole Population	1	1.63 - 2.23	C	1.93	0	0 - 0.3	C	0	0	0.9 - 1.5	C	1.2	0	1

None this event.

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.