C-Reactive Protein

Name	Line No.	Specimen 1 Range & Type	Specimen 1 Mean	Specimen 1 SD	Specimen 2 Range & Type	Specimen 2 Mean	Specimen 2 SD	Specimen 3 Range & Type	Specimen 3 Mean	Specimen 3 SD	Specimen 4 Range & Type	Specimen 4 Mean	Specimen 4 SD	Specimen 5 Range & Type	Specimen 5 Mean	Specimen 5 SD	No. of Labs
C-Reactive Protein, quantitative																	
Initial Grouping by Reagent																	
Siemens Dimension Flex	1	2.2 - 3.1	S 2.68	0.15	0 - 0.6	C 0.19	0.06	-			-			-			34
Beckman Olympus turb	2	2.3 - 3.1	S 2.69	0.14	0 - 0.9	S 0.24	0.22	-			-			-			21
Ortho Vitros	3	2.7 - 4.3	S 3.53	0.27	0.1 - 0.9	C 0.48	0.08	-			-			-			18
Abbott Aeroset/Architect	4	2.3 - 3.1	C 2.67	0.09	0 - 0.5	C 0.07	0.08	-			-			-			12
Initial Grouping by Sensitivity or Principle																	
All direct quantitive meth	5	1.5 - 4.1	S 2.8	0.44	0 - 0.8	S 0.22	0.18	-			-			-			91
All wide range methods	6	2.3 - 3.1	S 2.68	0.14	0 - 0.9	S 0.21	0.22	-			-			-			28
Total Population																	
Whole Population	7	1.5 - 4.0	S 2.75	0.42	0 - 0.8	S 0.22	0.2	-			-			-			124