



**PROFICIENCY TESTING SERVICE  
AMERICAN ASSOCIATION OF BIOANALYSTS**  
205 West Levee St. w Brownsville, TX 78520-5596  
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

**PARTICIPANT STATISTICS**

**THIRD QUADRIMESTER 2019**

**ORAL FLUIDS HIV-1 ANTIBODIES**

anti-HIV-1 Screening	Specimen 11			Specimen 12			Specimen 13			Specimen 14			Specimen 15		
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind
Avioq HIV-1		8		8			8			8					8
OraQuick ADVANCE Rapid		3		3			3			3					3
<b>Total Population</b>	<b>0</b>	<b>12</b>	<b>0</b>	<b>12</b>	<b>0</b>	<b>0</b>	<b>12</b>	<b>0</b>	<b>0</b>	<b>12</b>	<b>0</b>	<b>0</b>	<b>0</b>	<b>12</b>	<b>0</b>
<b>Flagging</b>	<b>***</b>		<b>***</b>		<b>&amp;&amp;&amp;</b>	<b>***</b>		<b>***</b>	<b>***</b>		<b>***</b>	<b>***</b>	<b>***</b>		<b>***</b>

& 4th Generation Ag/Ab Combo kits reacted to the p24 antigen as would be expected. They were evaluated against both antigen and antibody. If your kit reports serology and antigen independently, you must report antigen detection separately or be flagged.

anti-HIV-1 Confirmation	Specimen 11			Specimen 12			Specimen 13			Specimen 14			Specimen 15		
	Negative	Positive	Confirm testing not performed on neg	Negative	Positive	Confirm testing not performed on neg	Negative	Positive	Confirm testing not performed on neg	Negative	Indeterminate	Confirm testing not performed on neg	Negative	Positive	Confirm testing not performed on neg
In house validated		3		2		1	2		1	2		1			3
<b>Total Population</b>	<b>0</b>	<b>5</b>	<b>0</b>	<b>2</b>	<b>0</b>	<b>3</b>	<b>2</b>	<b>0</b>	<b>3</b>	<b>2</b>	<b>0</b>	<b>3</b>	<b>0</b>	<b>5</b>	<b>0</b>
<b>Flagging</b>	<b>***</b>		<b>***</b>		<b>***</b>		<b>***</b>		<b>***</b>		<b>***</b>		<b>***</b>		<b>***</b>

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.