



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. w Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2019

RUBELLA

Rubella IgG Method	Specimen 11		Specimen 12		Specimen 13		Specimen 14		Specimen 15		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere Rubella-Plus	7		1	6		7	7		7		7
ASI Rubella Test	12			12		12	12		12		12
Beckman Access IgG	10			10		10	10		10		10
Biokit Rbgen/Fisher SureVue	3			3		3	3		2	1	3
Bio-Rad Bioplex 2000, 2200	5			5		5	5		5		5
Fisher Sure-Vue	7			7		7	7		7		7
Roche Elecsys	4			4		4	4		4		4
Siemens Immulite 2000, 2500	6			6		6	6		6		6
Siemens Immuno 1/Advia IgG	3			3		3	3		3		3
Total Population	78	0	1	77	0	77	78	0	77	1	78
Flagging		***	***		***			***		***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.