



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2019

Shiga Toxin

Method	Specimen 11		Specimen 12		No. of Labs
	Neg	Pos	Neg	Pos	
Alere	4		4		4
Meridian ImmunoCard Stat	5		5		5
Total Population	9	0	9	0	9
Flagging		***		***	

Method	Specimen 11		Specimen 12		No. of Labs
	Neg	Pos	Neg	Pos	
Alere		4	4		4
Meridian ImmunoCard Stat		5	5		5
Total Population	0	9	9	0	9
Flagging	***			***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees.
 Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.