



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. w Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2019

STREP GROUP A ANTIGEN SCREEN

Strep A Screen Method	Specimen 11		Specimen 12		Specimen 13		Specimen 14		Specimen 15		No. of Labs
	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	Neg	Pos	
Alere BioSta Acceava wvd		3		3		3		3		3	3
Alere i		9		9		9		9		9	9
BD Veritor		6		6		6		6		6	6
Cardinal Health SP		3		3		3		3		3	3
Cepheid GeneXpert		8		8		8		8		8	8
Consult Dipstick, waived		6		6		6		6		6	6
Fisher Sure-Vue, waived		10		10		10		10		10	10
Henry Schein OneStep+, Wv	1	2		3		3		3		3	3
McKesson, waived		9		9		9		9		9	9
Meridian Diag ImmunoCard		3		3		3		3		3	3
Quidel QuickVue +		7		7		7		7		6	1
Quidel QuickVue Dipstick		10		10		10		10		10	10
Quidel QuickVue In-Line	1	33	4	30		34	3	31		34	34
Quidel Sofia		7		7		7		7		7	7
Sekisui OSOM		9		9		9		9		9	9
Sekisui OSOM Ultra		12		12		12		12		12	12
Sekisui OSOM, waived		21		21		21		21		21	21
Sekisui Rapid Strep A		7		7		7		7		7	7
Total Population	3	180	4	179	182	1	4	179	181	2	183
Flagging	***		***			***	***			***	

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.