



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. ♦ Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

THIRD QUADRIMESTER 2019

VIRAL ANTIGEN SCREEN, Waived

Influenza Type A	Specimen 11			Specimen 12		
	Neg	Pos	Pos for A or B	Neg	Pos	Pos for A or B
Alere	5	1		6		
BD Veritor	60	2	1	60	1	2
Binax NOW Flu A&B, wv	3			3		
BioSign	4			4		
Consult Diagnostics	11			11		
ID Now/Alere i	22			21		1
OraSure	3			3		
Quidel Quick Vue Infl A+B	9			9		
Quidel Sofia	29			30		
Sekisui OSOM A&B	11			11		
Total Population	162	3	1	162	1	4
Flagging		***	***		***	

Influenza Type B	Specimen 11		Specimen 12	
	Neg	Pos	Neg	Pos
ID Now/Alere i	20	1	1	20
Alere	5	1	1	5
BD Veritor	59	2	1	60
Binax NOW Flu A&B, wv	3			3
BioSign	4			4
Consult Diagnostics	11		1	10
OraSure	3			3
Quidel Quick Vue Infl A+B	10		7	3
Quidel Sofia	29		1	29
Sekisui OSOM A&B	11			11
Total Population	160	4	12	153
Flagging		***	***	

Respiratory Syncytial Virus	Specimen 11		Specimen 12	
	Neg	Pos	Neg	Pos
Alere Binax Now	20		20	
BD Veritor	8		8	
ID Now/Alere i	5		5	
Quidel QuickVue	9		9	
Quidel Sofia	6		6	
Total Population	50	0	50	0
Flagging		***		***

Correct responses are defined as those reflecting agreement among 80% or more of all participants or referees. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.