C-Reactive Protein

Name	Line No.	Specimen 1 Range & Type	Specimen 1 Mean	Specimen 1 SD	Specimen 2 Range & Type	Specimen 2 Mean	Specimen 2 SD	Specimen 3 Range & Type	Specimen 3 Mean	Specimen 3 SD	Specimen 4 Range & Type	Specimen 4 Mean	Specimen 4 SD	Specimen 5 Range & Type	Specimen 5 Mean	Specimen 5 SD	No. of Labs
C-Reactive Protein, quantitative																	
Initial Grouping by Reagent																	
Beckman Olympus turb	1	0 - 0.8	S 0.21	0.2	0 - 11.0	S 2.07	2.98	-			-			-			17
Siemens Dimension Flex	2	0 - 0.6	C 0.21	0.03	0.8 - 1.6	S 1.21	0.14	-			-			-			26
Ortho Vitros	3	0 - 1.0	S 0.41	0.19	0 - 20.0	S 3.86	5.39	-			-			-			13
Siemens CRP Extended RCRP	4	0 - 0.7	S 0.18	0.17	0.9 - 1.7	C 1.26	0.1	-			-			-			5
Roche Tina-Quant / CRPL3	5	0 - 0.5	C 0.1	0.08	0.7 - 1.5	C 1.1	0.09	-			-			-			9
Roche CRP High Sensitivity	6	0 - 0.5	C 0.05	0.09	0 - 181.8	S 30.4	50.46	-			-			-			5
Abbott Aeroset/Architect	7	0 - 0.5	C 0.06	0.07	0.9 - 1.7	C 1.28	0.06	-			-			-			9
Initial Grouping by Sensitivity or Principle																	
All wide range methods	8	0 - 0.8	S 0.19	0.19	0 - 9.9	S 2.0	2.64	-			-			-			23
All direct quantitive meth	9	0 - 0.7	S 0.2	0.17	0 - 45.7	S 3.52	14.06	-			-			-			74
Total Population																	
Whole Population	10	0 - 0.7	S 0.2	0.17	0 - 39.9	S 3.23	12.21	-			-			-			100