



## Hematology With Diff G

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	
<b>Leukocytes - Module G</b>																	
<b>Initial Grouping by Reagent</b>																	
Coulter Ac*T 5	1	5.8 - 7.9	P 6.84	0.31	2.1 - 2.8	P 2.43	0.17	13.7 - 18.5	P 16.07	0.43	21.3 - 28.8	P 25.07	0.49	4.2 - 5.6	P 4.91	0.1	13
Horiba ABX Pentra 60 C+	2	5.7 - 7.8	P 6.74	0.25	2.0 - 2.8	P 2.41	0.08	13.4 - 18.1	P 15.74	0.51	21.1 - 28.6	P 24.84	0.93	4.0 - 5.5	P 4.74	0.13	15
Horiba ABX Pentra 80	3	5.8 - 7.9	P 6.86	0.16	2.1 - 2.9	P 2.51	0.08	13.4 - 18.2	P 15.8	0.36	20.9 - 28.3	P 24.61	0.81	4.1 - 5.6	P 4.84	0.14	16
<b>Initial Grouping by Sensitivity or Principle</b>																	
Pentra 60 / Ac*T 5	4	5.8 - 7.8	P 6.78	0.27	2.1 - 2.8	P 2.42	0.12	13.5 - 18.3	P 15.88	0.49	21.2 - 28.7	P 24.99	0.76	4.1 - 5.5	P 4.81	0.14	30
Other Instruments	5	5.8 - 7.9	P 6.86	0.16	2.1 - 2.9	P 2.51	0.08	13.4 - 18.2	P 15.8	0.36	20.9 - 28.3	P 24.61	0.81	4.1 - 5.6	P 4.84	0.14	16
<b>Total Population</b>																	
Whole Population	6	5.8 - 7.8	P 6.82	0.25	2.1 - 2.8	P 2.45	0.12	13.5 - 18.3	P 15.89	0.5	21.2 - 28.7	P 24.92	0.87	4.1 - 5.5	P 4.82	0.15	49
<b>Erythrocytes - Module G</b>																	
<b>Initial Grouping by Reagent</b>																	
Coulter Ac*T 5	1	4.38 - 4.94	P 4.663	0.071	1.91 - 2.16	P 2.036	0.035	5.22 - 5.89	P 5.553	0.087	2.92 - 3.29	P 3.103	0.054	4.04 - 4.56	P 4.302	0.068	13
Horiba ABX Pentra 60 C+	2	4.32 - 4.87	P 4.594	0.087	1.9 - 2.14	P 2.021	0.021	5.14 - 5.8	P 5.471	0.092	2.9 - 3.27	P 3.089	0.051	4.0 - 4.52	P 4.26	0.073	14
Horiba ABX Pentra 80	3	4.32 - 4.87	P 4.594	0.074	1.85 - 2.08	P 1.966	0.042	5.18 - 5.84	P 5.506	0.097	2.85 - 3.21	P 3.031	0.066	3.98 - 4.49	P 4.236	0.067	16
<b>Initial Grouping by Sensitivity or Principle</b>																	
Pentra 60 / Ac*T 5	4	4.34 - 4.9	P 4.62	0.086	1.9 - 2.15	P 2.026	0.031	5.17 - 5.84	P 5.505	0.098	2.91 - 3.28	P 3.093	0.055	4.02 - 4.53	P 4.274	0.076	29
Other Instruments	5	4.32 - 4.87	P 4.594	0.074	1.85 - 2.08	P 1.966	0.042	5.18 - 5.84	P 5.506	0.097	2.85 - 3.21	P 3.031	0.066	3.98 - 4.49	P 4.236	0.067	16
<b>Total Population</b>																	
Whole Population	6	4.33 - 4.89	P 4.61	0.081	1.89 - 2.13	P 2.006	0.045	5.17 - 5.83	P 5.504	0.097	2.89 - 3.26	P 3.072	0.065	4.01 - 4.52	P 4.263	0.074	49
<b>Hemoglobin - Module G</b>																	
<b>Initial Grouping by Reagent</b>																	
Coulter Ac*T 5	1	12.5 - 14.4	P 13.48	0.07	6.2 - 7.2	P 6.72	0.09	15.6 - 17.9	P 16.74	0.14	7.6 - 8.8	P 8.19	0.06	11.4 - 13.2	P 12.31	0.09	13
Horiba ABX Pentra 60 C+	2	12.5 - 14.4	P 13.49	0.17	6.1 - 7.1	P 6.59	0.09	15.7 - 18.1	P 16.91	0.22	7.5 - 8.7	P 8.09	0.09	11.4 - 13.1	P 12.26	0.17	15
Horiba ABX Pentra 80	3	12.5 - 14.4	P 13.49	0.17	6.1 - 7.0	P 6.56	0.11	15.8 - 18.2	P 16.98	0.18	7.5 - 8.6	P 8.06	0.11	11.4 - 13.1	P 12.28	0.14	16
<b>Initial Grouping by Sensitivity or Principle</b>																	
Pentra 60 / Ac*T 5	4	12.5 - 14.4	P 13.48	0.14	6.2 - 7.1	P 6.63	0.12	15.7 - 18.0	P 16.85	0.23	7.6 - 8.7	P 8.13	0.11	11.4 - 13.1	P 12.27	0.16	30
Other Instruments	5	12.5 - 14.4	P 13.49	0.17	6.1 - 7.0	P 6.56	0.11	15.8 - 18.2	P 16.98	0.18	7.5 - 8.6	P 8.06	0.11	11.4 - 13.1	P 12.28	0.14	16
<b>Total Population</b>																	
Whole Population	6	12.5 - 14.4	P 13.49	0.15	6.1 - 7.1	P 6.61	0.12	15.7 - 18.1	P 16.89	0.22	7.5 - 8.7	P 8.1	0.12	11.4 - 13.1	P 12.28	0.15	49
<b>Hematocrit - Module G</b>																	
<b>Initial Grouping by Reagent</b>																	
Coulter Ac*T 5	1	36.0 - 40.7	P 38.35	0.51	17.3 - 19.5	P 18.42	0.56	44.9 - 50.6	P 47.78	0.82	22.3 - 25.1	P 23.68	0.58	32.7 - 36.9	P 34.83	0.8	13
Horiba ABX Pentra 60 C+	2	36.4 - 41.0	P 38.69	0.82	17.4 - 19.6	P 18.47	0.25	45.4 - 51.2	P 48.27	1.09	22.5 - 25.3	P 23.89	0.52	33.0 - 37.2	P 35.14	0.73	15
Horiba ABX Pentra 80	3	36.3 - 41.0	P 38.65	0.75	17.6 - 19.9	P 18.77	0.46	45.1 - 50.9	P 48.01	0.98	22.6 - 25.5	P 24.04	0.46	33.0 - 37.2	P 35.14	0.55	16
<b>Initial Grouping by Sensitivity or Principle</b>																	
Pentra 60 / Ac*T 5	4	36.2 - 40.8	P 38.52	0.71	17.3 - 19.5	P 18.43	0.42	45.2 - 50.9	P 48.04	0.99	22.4 - 25.2	P 23.78	0.56	32.9 - 37.1	P 34.97	0.78	30
Other Instruments	5	36.3 - 41.0	P 38.65	0.75	17.6 - 19.9	P 18.77	0.46	45.1 - 50.9	P 48.01	0.98	22.6 - 25.5	P 24.04	0.46	33.0 - 37.2	P 35.14	0.55	16
<b>Total Population</b>																	
Whole Population	6	36.3 - 40.9	P 38.57	0.71	17.4 - 19.7	P 18.55	0.46	45.2 - 50.9	P 48.04	0.97	22.4 - 25.3	P 23.86	0.53	32.9 - 37.2	P 35.05	0.71	49
<b>Platelets - Module G</b>																	
<b>Initial Grouping by Reagent</b>																	
Coulter Ac*T 5	1	189 - 315	P 252.3	6.3	55 - 91	P 73.1	5.1	225 - 374	P 299.4	10.1	300 - 500	P 400.3	12.4	178 - 297	P 237.7	9.3	13
Horiba ABX Pentra 60 C+	2	188 - 313	P 250.0	8.7	53 - 88	P 70.6	4.7	224 - 374	P 299.2	14.0	302 - 503	P 402.5	18.5	175 - 292	P 233.7	11.2	15

Horiba ABX Pentra 80	3	189 - 315	P 251.7	8.3	51 - 85	P 68.4	4.6	227 - 378	P 302.3	10.7	299 - 498	P 398.1	14.7	175 - 292	P 233.6	9.6	16
<b>Initial Grouping bySensitivityor Principle</b>																	
Pentra 60 / Ac*T 5	4	188 - 314	P 251.1	7.5	54 - 90	P 72.0	5.1	225 - 375	P 299.6	11.8	301 - 501	P 401.2	15.9	176 - 294	P 235.3	10.3	30
Other Instruments	5	189 - 315	P 251.7	8.3	51 - 85	P 68.4	4.6	227 - 378	P 302.3	10.7	299 - 498	P 398.1	14.7	175 - 292	P 233.6	9.6	16
<b>Total Population</b>																	
Whole Population	6	189 - 314	P 251.4	7.7	53 - 89	P 70.8	5.2	225 - 375	P 300.4	11.3	300 - 500	P 400.0	15.3	176 - 294	P 235.0	9.9	49

### Neutrophil % - Module G

<b>Initial Grouping byReagent</b>																	
Coulter Ac*T 5	1	47.6 - 67.5	S 57.52	3.32	49.3 - 64.4	S 56.81	2.52	41.3 - 61.2	S 51.26	3.31	45.7 - 70.5	S 58.11	4.13	45.8 - 64.0	S 54.87	3.03	13
Horiba ABX Pentra 60 C+	2	48.5 - 75.6	S 62.09	4.52	45.6 - 71.9	S 58.75	4.37	47.2 - 64.9	S 56.07	2.96	52.3 - 74.3	S 63.28	3.67	50.1 - 68.4	S 59.22	3.05	15
Horiba ABX Pentra 80	3	48.1 - 72.6	S 60.33	4.09	43.9 - 77.2	S 60.57	5.55	37.8 - 65.2	S 51.51	4.58	48.0 - 75.7	S 61.83	4.62	47.5 - 71.2	S 59.33	3.95	16
<b>Initial Grouping bySensitivityor Principle</b>																	
Pentra 60 / Ac*T 5	4	47.0 - 73.6	S 60.31	4.44	46.7 - 69.6	S 58.13	3.81	42.5 - 65.3	S 53.89	3.8	47.0 - 76.0	S 61.53	4.83	46.1 - 68.7	S 57.41	3.76	30
Other Instruments	5	48.1 - 72.6	S 60.33	4.09	43.9 - 77.2	S 60.57	5.55	37.8 - 65.2	S 51.51	4.58	48.0 - 75.7	S 61.83	4.62	47.5 - 71.2	S 59.33	3.95	16
<b>Total Population</b>																	
Whole Population	6	47.4 - 73.7	S 60.55	4.37	45.0 - 73.1	S 59.08	4.69	40.5 - 65.8	S 53.15	4.21	47.6 - 76.0	S 61.81	4.75	46.5 - 69.7	S 58.13	3.86	49

### Lymphocyte % - Module G

<b>Initial Grouping byReagent</b>																	
Coulter Ac*T 5	1	21.0 - 34.3	S 27.63	2.22	10.7 - 32.4	S 21.58	3.61	28.0 - 47.3	S 37.69	3.22	13.2 - 35.9	S 24.58	3.78	17.4 - 36.4	S 26.93	3.17	13
Horiba ABX Pentra 60 C+	2	14.2 - 35.9	S 25.09	3.62	11.4 - 27.6	S 19.5	2.69	23.1 - 42.7	S 32.91	3.26	15.4 - 27.9	S 21.64	2.09	14.8 - 30.0	S 22.4	2.52	15
Horiba ABX Pentra 80	3	18.2 - 42.0	S 30.08	3.97	11.7 - 37.0	S 24.36	4.21	27.4 - 54.1	S 40.78	4.44	18.4 - 43.0	S 30.68	4.11	18.1 - 36.6	S 27.35	3.08	16
<b>Initial Grouping bySensitivityor Principle</b>																	
Pentra 60 / Ac*T 5	4	15.6 - 37.3	S 26.46	3.62	10.5 - 31.0	S 20.78	3.41	22.8 - 48.0	S 35.41	4.21	13.3 - 33.0	S 23.13	3.28	13.1 - 36.5	S 24.8	3.91	30
Other Instruments	5	18.2 - 42.0	S 30.08	3.97	11.7 - 37.0	S 24.36	4.21	27.4 - 54.1	S 40.78	4.44	18.4 - 43.0	S 30.68	4.11	18.1 - 36.6	S 27.35	3.08	16
<b>Total Population</b>																	
Whole Population	6	14.7 - 40.3	S 27.49	4.26	9.7 - 33.9	S 21.81	4.02	22.3 - 52.1	S 37.23	4.96	9.9 - 41.0	S 25.45	5.18	14.3 - 37.1	S 25.73	3.8	49

### Monocyte % - Module G

<b>Initial Grouping byReagent</b>																	
Horiba ABX Pentra 60 C+	1	0 - 2.1	S 0.81	0.44	0 - 1.8	S 0.46	0.44	0 - 3.6	S 1.43	0.74	0 - 2.8	S 1.38	0.46	0 - 1.9	S 0.61	0.44	15
Coulter Ac*T 5	2	0 - 4.3	S 1.44	0.95	0 - 1.6	C 0.62	0.28	0 - 3.7	S 1.49	0.73	0 - 4.0	S 1.58	0.82	0 - 2.4	S 1.11	0.44	13
Horiba ABX Pentra 80	3	0 - 2.5	S 0.83	0.54	0 - 1.3	C 0.29	0.23	0 - 1.8	S 0.66	0.36	0 - 2.4	S 1.04	0.45	0 - 1.5	S 0.41	0.35	16
<b>Initial Grouping bySensitivityor Principle</b>																	
Pentra 60 / Ac*T 5	4	0 - 3.3	S 1.06	0.76	0 - 1.7	S 0.53	0.39	0 - 3.6	S 1.47	0.73	0 - 3.4	S 1.51	0.63	0 - 2.4	S 0.87	0.52	30
Other Instruments	5	0 - 2.5	S 0.83	0.54	0 - 1.3	C 0.29	0.23	0 - 1.8	S 0.66	0.36	0 - 2.4	S 1.04	0.45	0 - 1.5	S 0.41	0.35	16
<b>Total Population</b>																	
Whole Population	6	0 - 3.0	S 0.98	0.68	0 - 1.5	S 0.46	0.36	0 - 3.3	S 1.17	0.73	0 - 3.2	S 1.34	0.61	0 - 2.2	S 0.7	0.51	49

### Eosinophil % - Module G

<b>Initial Grouping byReagent</b>																	
Horiba ABX Pentra 60 C+	1	2.7 - 9.6	S 6.15	1.15	2.3 - 17.5	S 9.89	2.54	3.4 - 8.7	S 6.09	0.88	2.8 - 5.5	S 4.17	0.44	2.7 - 16.7	S 9.73	2.33	15
Coulter Ac*T 5	2	2.8 - 9.6	S 6.19	1.12	0.7 - 17.5	S 9.11	2.79	1.6 - 8.9	S 5.24	1.21	1.7 - 5.8	S 3.74	0.7	3.1 - 14.7	S 8.88	1.93	13
Horiba ABX Pentra 80	3	3.2 - 11.1	S 7.13	1.33	3.3 - 18.6	S 10.94	2.54	2.0 - 10.0	S 5.99	1.33	2.1 - 5.2	S 3.64	0.52	5.2 - 15.6	S 10.39	1.74	16
<b>Initial Grouping bySensitivityor Principle</b>																	
Pentra 60 / Ac*T 5	4	2.9 - 9.5	S 6.19	1.11	1.8 - 17.4	S 9.64	2.6	2.3 - 9.0	S 5.64	1.12	1.9 - 5.9	S 3.9	0.66	3.0 - 15.7	S 9.32	2.12	30
Other Instruments	5	3.2 - 11.1	S 7.13	1.33	3.3 - 18.6	S 10.94	2.54	2.0 - 10.0	S 5.99	1.33	2.1 - 5.2	S 3.64	0.52	5.2 - 15.6	S 10.39	1.74	16
<b>Total Population</b>																	
Whole Population	6	2.6 - 10.4	S 6.49	1.29	1.8 - 18.1	S 9.95	2.71	2.0 - 9.4	S 5.72	1.23	1.9 - 5.7	S 3.8	0.62	3.0 - 16.1	S 9.52	2.19	49

### Basophil % - Module G

<b>Initial Grouping byReagent</b>																	
Horiba ABX Pentra 60 C+	1	0 - 12.3	S 5.69	2.19	0 - 24.6	S 11.4	4.41	0 - 7.5	S 3.5	1.33	0 - 20.4	S 9.53	3.61	0 - 16.4	S 7.61	2.92	15
Coulter Ac*T 5	2	4.4 - 8.5	S 6.45	0.67	9.2 - 14.3	S 11.78	0.85	2.9 - 4.9	C 3.91	0.21	9.2 - 12.3	S 10.75	0.52	5.1 - 11.1	S 8.13	0.99	13
Horiba ABX Pentra 80	3	0 - 8.6	S 1.63	2.31	0 - 15.2	S 2.94	4.1	0 - 5.6	S 1.07	1.51	0 - 14.9	S 2.82	4.04	0 - 12.2	S 2.52	3.22	16
<b>Initial Grouping bySensitivityor Principle</b>																	
Pentra 60 / Ac*T 5	4	0 - 12.2	S 5.65	2.18	0 - 23.3	S 10.88	4.14	0 - 7.3	S 3.44	1.3	0 - 20.0	S 9.48	3.51	0 - 15.8	S 7.37	2.82	30
Other Instruments	5	0 - 8.6	S 1.63	2.31	0 - 15.2	S 2.94	4.1	0 - 5.6	S 1.07	1.51	0 - 14.9	S 2.82	4.04	0 - 12.2	S 2.52	3.22	16
<b>Total Population</b>																	

---

Whole Population	6	0 - 13.0	S 4.33	2.9	0 - 24.9	S 8.33	5.54	0 - 7.9	S 2.66	1.76	0 - 21.8	S 7.35	4.81	0 - 16.9	S 5.8	3.69	49
------------------	---	----------	--------	-----	----------	--------	------	---------	--------	------	----------	--------	------	----------	-------	------	----